


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AVH-5-N-A**

S.N.: **751**

Date: **FEB 24 82**



a) Output signal amplitude:
 ≥ 0 TO -15 V

b) Pulse width:
 ≤ 500 PSEC

c) Jitter, stability:
OK

d) Prime power:
 $+15 \pm 1.0$ VOLT

180 mA
(MAX)

10 VOLTS/DIV.

1.0 NSEC/DIV

400 KHZ.