


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVR-1-PW-0-1**

S.N.: **753**

Date: **APRIL 12 82**



- a) Output signal amplitude:
0 TO +300V
- b) Pulse width:
0.1 TO 5.0 μ sec
- c) Rise time:
 ≈ 15 nsec
- d) Fall time:
 ≈ 15 nsec.
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
110/220V
50-60 Hz.

50V / DIV
0.1 μ sec / DIV
 $R_L = 50\Omega$