


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AUG-1-P-A

S.N.: 763

Date: FEB 3 82



- a) Output signal amplitude:
0 to +250 V
- b) Pulse width:
≤ 4 ns (2.5 divisions)
- c) Jitter, stability:
OK
- d) Prime power:

+15 ± 1.0 V

325 mA

(max.)

100 V/DIV
5.0 ns/ER/DIV
40 kHz
 $R_i = 50 \Omega$