

PULSE GENERATOR
PERFORMANCE CHECK

Model:


PV0-2-A

S.N.:

766

Date:

MAR 21 82



- a) Output signal amplitude:
0 TO +10 AMPS
- b) Pulse width:
0 TO 15 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT

275 mA (MAX)

7 AMPS/DIV

5 NSEC/DIV

10 KHZ.

$R_L \approx 5 \Omega$.