


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AUL-TN-1**

S.N.: **773**

Date: **MAR 18 82**



100 VOLTS/DIV
5 NSEC/DIV

$R_L = 50 \Omega$

$C_L = 0$

1 KHz

a) Output signal amplitude:

- 230 V

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

≤ 1 NSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 5 KHz

f) Jitter, stability:

OK

g) Prime power:

110/220 V

50-60 Hz.

h) PROP DELAY:

95-100 NSEC.

i) OFFSET:

+30 VOLTS

OR 0 VOLTS.

[Signature]