


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-13*

S.N.: *780*

Date: *FEB 16 82*



a) Output signal amplitude:

$\approx + 30 \text{ VOLTS}$

b) Pulse width:

$\leq 0.5 \text{ NSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+ 15 \pm 1.0 \text{ VOLT}$

150 mA.

16 VOLTS/DIV

1.0 NSEC/DIV

400 KHz.