


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AUP-AV-2-05-PS-TTL-17**

S.N.: **789**

Date: **MAR 31 82**



- a) Output signal amplitude:
0 TO +10 V
- b) Pulse width:
5 TO 50 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
2.0 NSEC/DIV
40 KHZ

110/220 V
50 - 60 HZ.

*** UNIT SHIPPED**
WITH VOLTAGE
SELECTOR (MVC)
IN 220V POSITION

AY