

PULSE GENERATOR

PERFORMANCE CHECK

Model:


*AUP-AU-1-P-TTL-03-PS*

S.N.:


*792*

Date:

*MAR 31 82*



*10 V OUT/DIV  
1 NS/DIV*



*10 VOLTS/DIV  
1 NS/DIV  
400 KHz*

a) Output signal amplitude:

*0 TO +10 V*

b) Pulse width:

*0 TO 4 NS*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 1 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*110/220 V*

*50-60 Hz*