


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AUH-5-1-P-PS-A**

S.N.: **793**

Date: **MAR 18 82**



a) Output signal amplitude:

0 TO +10V

b) Pulse width:

≤ 135 nsec

c) Jitter, stability:

OK

d) Prime power:

120/240 V

50-60 Hz.

UNIT SHIPPED

WITH VOLTAGE

SET WITH (A) IN

240 V

POSITION.

10 VOLTS/DIV

1 nsec/DIV

400 kHz.