


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-P*

S.N.: *803*

Date: *APRIL 30 82*



*10 VOLTS/DIV  
1.0 NSEC/DIV  
4.0 MHz*

a) Output signal amplitude:

*0 TO +15V*

b) Pulse width:

*0 TO 2 NSEC*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 1.0 VOLT*

*450 mA (MAX)*