


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-S-A-1P*

S.N.: *804*

Date: *APRIL 30 82*



- a) Output signal amplitude:
 $\approx 20V_{max} (0 \rightarrow +20V)$
- b) Pulse width:
0.5 μ SEC
- c) Jitter, stability:
OK
- d) Prime power:

+15 \pm 1.0 VOLT.

130 mA.

[Signature]

20 VOLTS/DIV

1.0 μ SEC/DIV

400 KHZ.