


PULSE GENERATOR

PERFORMANCE CHECK


Model: **AUP-AV-2-N-TTL-03-PS-6A-6W**

S.N.: **809**

Date: **MAR 31 82**



**10 NS/DIV**  
**10 VOLTS/DIV**  
**40 KHZ**



**2 NSEC/DIV**  
**10 VOLTS/DIV**  
**40 KHZ.**

a) Output signal amplitude:

**0 TO -10V**

b) Pulse width:

**5 NS TO 50 NS**

c) Rise time:

**≤ 100 PSEC**

d) Fall time:

**≤ 135 PSEC**

e) PRF:

**0 TO 50 KHZ**

f) Jitter, stability:

**OK**

g) Prime power:

**110/220 V**

**50 - 60 HZ.**