


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5

S.N.: 810

Date: APRIL 30 82



a) Output signal amplitude:

$\approx 25 \text{ VOLTS}$

b) Pulse width:

$\leq 0.5 \text{ NSEC}$


c) Jitter, stability:

OK

d) Prime power:

$\pm 15 \pm 1.0 \text{ V}$

150 mA



20 VOLTS/DIV.

1.0 NSEC/DIV

400 KHZ.