


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AGM-2-N*
S.N.: *812*
Date: *APRIL 30 82*



- a) Output signal amplitude:
0 TO -15 V
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
- g) Prime power: *OK*

10 VOLTS/DIV.
1.0 NSEC/DIV
4.0 MHz.

+24 V
450 mA MAX