


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AU-1-P-05*


S.N.: *814*

Date: *MMAR 9 82*



*1.0 NSEC/DIV
10 VOLTS/DIV
400 KHZ.*

- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0.2 TO 4 NSEC
- c) Rise time:
≤ 100 NSEC
- d) Fall time:
≈ 135 NSEC
- e) PRF:
0 TO 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:



*1.0 NSEC/DIV
10 VOLTS/DIV
40 KHZ.*

*+15 ± 1.0 VOLT.
300 mA (MAX)*

[Signature]