


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH P
S.N.: 822
Date: MAY 24 82



- a) Output signal amplitude:
 $\geq +30 \text{ V}$
- b) Pulse width:
 $\leq 1 \text{ NSEC}$
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

16 VOLTS/DIV.
1.0 NSEC/DIV.
400 KHz.

$+15 \pm 1.0 \text{ V}$
135 mA (max)