


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-A-PS-P
S.N.: 839
Date: APRIL 30 82



- a) Output signal amplitude:
0 TO +50V (70Ω_{in})
- b) Pulse width:
0 TO 100 NSEC.
- c) Rise time:
≤ 0.5 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 32 VOLTS/DIV
10 NSEC/DIV
20 KHz.

110/220V
50-60 Hz