


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-D-PS-F*

S.N.: *840*

Date: *APRIL 30 82*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 2 NSEC.
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 135 PSEC.
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

110/220V
50-60 Hz.

10 VOLT/DIV
1.0 NSEC/DIV
4.0 MHz.