


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AUM-2-P-PS*

S.N.: *841*

Date: *MAY 13 82*



- a) Output signal amplitude:  
*0 TO +15*
- b) Pulse width:  
*0 TO 2 NSEC.*
- c) Rise time:  
*≤ 100 PSEC.*
- d) Fall time:  
*≤ 135 PSEC.*
- e) PRF:  
*0 TO 25 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*110/220 V*  
*50 - 60 MHz.*

*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*4.0 MHz.*