


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVL-AU-N-T-05-CCRS**
S.N.: **842**
Date: **APRIL 22 82**



- a) Output signal amplitude:
0 TO - 250V
- b) Pulse width:
0 TO 100 NS
- c) Rise time:
≤ 5 NS - 15 NS
- d) Fall time:
≤ 5 NS - 15 NS
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+ 24V DC
350 mA (MAX)

10 NS/DIV
100 VOLTS/DIV
2 KHZ
R_L = 50 Ω

NOTE UNIT REQUIRES
50Ω LOAD AT
ALL TIMES.