


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-3A-F
S.N.: 847
Date: APRIL 30 82



- a) Output signal amplitude:
+125 TO +75 AMPS
- b) Pulse width:
2 ± 0.5 NSEC
- c) Rise time: CHALF POWER
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

20 AMPS/DIV
2 NSEC/DIV
R_L = 5.1 Ω.
10 KHZ.

+15 ± 1.0 VOLTS
150 mA