


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AU-1-TTL-N*

S.N.: *852*

Date: *MAY 24 82*



10 VOLTS/DIV
1.0 NSEC/DIV
400 KHz

- a) Output signal amplitude:
0 TO -10 V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 10 V

275 mA

A large, stylized handwritten signature or scribble in the bottom right corner of the page.