


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-1*

S.N.: *854*

Date: *MAY 4 1982*



- a) Output signal amplitude:
25 TO 100 AMPS.
- b) Pulse width:
25 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHZ.
- f) Jitter, stability:
OK
- g) Prime power:

*64 AMPS/DIV.
10 NSEC/DIV.
1 KHZ.*

*+15 ± 1.0 VOLTS.
150 mA.*