


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2-N

S.N.: 858

Date: MAY 4 82



- a) Output signal amplitude:
0 TO -10 AMPS
- b) Pulse width:
0 TO 15 NSEC.
- c) Rise time:
1 NSEC
- d) Fall time:
1 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

5 NSEC/DIV.

20 AMPS/DIV.

$R_L = 5-1 \Omega$

PRF = 20 KHZ.

$+15 \pm 1.0 \text{ VOLT}$

150 mA

[Handwritten signature]