


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-5-1-A7*

S.N.: *862*

Date: *MAY 24 82*



- a) Output signal amplitude:  
*0 TO +10V*
- b) Pulse width:  
*≤ 135 PSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 10 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 Volts/DIV.  
1.0 NSEC/DIV.  
400 KHz.*

*+15V ± 1.0V  
160 mA (MAX)*

*[Signature]*