


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-N-PS

S.N.: 864

Date: JUNE 23 82



320V/DIV

10 NSEC/DIV

2 KHZ

a) Output signal amplitude:

0 TO - 350V

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

~ 1.5 NSEC

d) Fall time:

~ 2 NSEC

e) PRF:

0 TO 5 KHZ

f) Jitter, stability:

OK

g) Prime power:

110/220V,
50-60 HZ