


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-V-N-105-A


S.N.: 867

Date: June 23 82



32 V/DIV
10 NSEC/DIV
10 KHZ

- a) Output signal amplitude:
0 TO -50V
- b) Pulse width:
0 TO 100 NSEC.
- c) Rise time:
≤ 0.5 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
110/220V,
50-60 HZ.



RISE TIME
32 V/DIV
1.0 NSEC/DIV
10 KHZ..