

PULSE GENERATOR  
PERFORMANCE CHECK

Model:


MVI-U-N-A-PS

S.N.:


868

Date:

Jan 23 82.



32 V/DIV.  
10 NSEC/DIV.  
10 KHz



RISE TIME  
32 V/DIV  
1.0 NSEC/DIV  
10 KHz.

a) Output signal amplitude:

0 TO 50V.

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

≤ 0.5 NSEC.

d) Fall time:

≤ 2 NSEC.

e) PRF:

0 TO 20 KHz

f) Jitter, stability:

OK

g) Prime power:

110/220V,  
50-60 Hz.