

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVL-2-P

S.N.:

877

Date:

JUNE 28 82



a) Output signal amplitude:

0 TO ± 350 V.

b) Pulse width:

0 TO 100 NSOC.

c) Rise time:

≤ 1.5 NSEC

d) Fall time:

≤ 2 NSEC.

e) PRF:

0 TO 5 KHZ.

f) Jitter, stability:

OK

g) Prime power:

$\pm 15 \pm 1.0$ V

500 mA (max)