


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVO-P**

S.N.: **879**

Date: **JUNE 28 82**



- a) Output signal amplitude:  
**10-50 AMPs**
- b) Pulse width:  
**≈ 25 NSs**
- c) Rise time:  
**NA**
- d) Fall time:  
**NA**
- e) PRF:  
**0 TO 1KHz**
- f) Jitter, stability:  
**OK**
- g) Prime power:

**20 AMPs/DIV**  
**10 NSs/DIV**  
**1 KHz**

**+15 ± 1.0 V**  
**180 mA**  
**(MAX)**  
**[Signature]**