


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVI-V-P**
S.N.: **885**
Date: **July 23 82**



32 V/DIV
10 nsec/DIV.

- a) Output signal amplitude:
± 50 V
- b) Pulse width:
0 TO 100 nsec.
- c) Rise time:
≤ 0.5 nsec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V
175 mA.

ALY