


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-5-1-P  
S.N.: 887  
Date: MAY 24 82



- a) Output signal amplitude:  
 $\geq +10V$
- b) Pulse width:  
 $\leq 135 \text{ PSEC}$
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 MHz.
- f) Jitter, stability:  
OK
- g) Prime power:

10 VOLTS/DIV.  
1.0 NSEC/DIV  
400 KHz

+15V  $\pm$  1.0V  
150 mA (MAX)