


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-2-1^{*}

S.N.: 892

Date: MAY 24 82



- a) Output signal amplitude:
 $\geq +20V$
- b) Pulse width:
 $\leq 0.5 \mu\text{SEC}$
- c) Rise time:
NA
- d) Fall time:
NA
- e) * PRF:
0 TO 2.5 MHz
- f) Jitter, stability:
OK
- g) Prime power:

$+15 \pm 1.0V$
180 mA

10 VOLTS/DIV
1.0 $\mu\text{SEC/DIV}$
2.0 MHz

* AVH-5 MODIFIED
FOR 2 MHz
PRF OPERATION.