


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-2-1*
S.N.: 893
Date: MARCH 24 82



a) Output signal amplitude:

$\geq +20V$

b) Pulse width:

$\leq 0.5 \text{ nsec}$

c) Rise time:

NA

d) Fall time:

NA

e) * PRF:

0 TO 2.5 MHz

f) Jitter, stability:

OK

g) Prime power:

$+15 \pm 1.0 \text{ VOLTS}$

170 mA.

10 VOLTS/DIV

1.0 nsec/DIV

2.0 MHz

* AVH-5 MODIFIED
FOR 2 MHz PRF
OPERATION.