


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AU-1-N-05-C*

S.N.: *894*

Date: *MAY 31 82*



- a) Output signal amplitude:
0 TO 10V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:^{*}
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

110/220V, 50-60 Hz

** 1 kHz TO 1 MHz ON INTERNAL CLOCK.*

100 NSEC/DIV.

10 VOLTS/DIV

*400 kHz
INTERNAL CLOCK.*