


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVL-2-PS-03-P**

S.N.: **897**

Date: **July 16 82**



- a) Output signal amplitude:
0 to +350 V
- b) Pulse width:
0 to 100 nsec.
- c) Rise time:
≤ 1 nsec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

**320 VOLTS/DIV.
10 nsec/DIV.
1 kHz.**

**110/220 V
50 - 60 Hz.**