


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-N-D*

S.N.: *910*

Date: *JUN 28 82*



a) Output signal amplitude:

0 TO -15V

b) Pulse width:

0 TO 2 NS

c) Rise time:

≤ 100 PSEZ

d) Fall time:

≤ 135 PSEZ

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 V

450 mA

1.0 NS/25V/DIV

10 V/25V/DIV

4.0 MHz

(MTH)
[Signature]