


PULSE GENERATOR
PERFORMANCE CHECK

Model: APV-19U-2-N-5TL

S.N.: 913

Date: JUN 22 82



- a) Output signal amplitude:
0 TO 10V
- b) Pulse width:
5 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

5 NSEC/DIV.
10 VOLTS/DIV
40 KHZ.

+15 ± 1.0 V
225 mA (MAX)