


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-P*  
S.N.: *924*  
Date: *MAY 31 82*



- a) Output signal amplitude:  
*0 TO +5V*
- b) Pulse width:  
*0 TO 6 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 150 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5.0 VOLT/DIV*  
*1.0 NSEC/DIV*  
*4.0 MHz*

*+24 ± 1.0 VOLT*

*380 mA (MAX)*