


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT-2-N*

S.N.: *935*

Date: *Jul 31 82*



a) Output signal amplitude:

*≤ -60 V*

b) Pulse width:

*≤ 2 nsec*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0V*

*160 mA*

*[Signature]*

*32 VOLTS/DIV.*

*2 nsec/DIV*

*400 kHz.*