


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **A20-1-0P2 + A20-1-0P25**

S.N.: **940 d 941**

Date: **AUG 11 82**



a) Output signal amplitude:

+100 AMPS MAX

b) Pulse width:

6-10 NSEC (H.P.)

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V,

350 mA (MAX)

e) PRF: **DC - 5 KHZ**

100 AMPS/DIV

10 NSEC/DIV.

4 KHZ.

$R_L = 1 \Omega$