


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AV14-5-1-P-1A

S.N.: 943

Date: AUG 22 82



a) Output signal amplitude:

0.6 ± 10 V

b) Pulse width:

≈ 130 psec

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V

175 mA rms

10 VOLTS/DIV

1.0 nsec/DIV

400 kHz