


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-2-P-A-PS

S.N.: 945

Date: 1946 6 72



100 V / DIV.
2 NSEC / DIV
40 KHZ.

a) Output signal amplitude:

0 TO +125V

b) Pulse width:

≤ 2 NSEC

c) Jitter, stability:

OK

d) Prime power:

120/240 V,
50 - 60 HZ.

NOTE: UNIT
SHIPPED
WITH VOLTAGE
SELECTOR
CARD IN
120 V
POSITION.