


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO -1*

S.N.: *947*

Date: *JUNE 28 82*



- a) Output signal amplitude:
25 TO 100 AMPS.
- b) Pulse width:
= 25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1KHZ
- f) Jitter, stability:
OK.
- g) Prime power:

*67 AMPS/DIV.
10 NSEC/DIV
1 KHZ*

*+15 ± 1.0V
160 mA*