

PULSE GENERATOR

PERFORMANCE CHECK

Model:


A40-1

S.N.:

949

Date:

JUNE 28 82



a) Output signal amplitude:

25 TO 100 AMPS

b) Pulse width:

~ 25 NSEC.

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 KHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

145 mA.

67 AMPS/DIV

10 NSEC/DIV

1 KHz.