


PULSE GENERATOR  
PERFORMANCE CHECK

Model: **AVK**  
S.N.: **952**  
Date: **Jul 31 82**



- a) Output signal amplitude:  
**> +5 VOLTS**
- b) Pulse width:  
**4 NSEC - 100 NSEC**
- c) Rise time:  
**≤ 100 PSEC**
- d) Fall time:  
**≤ 2 NSEC.**
- e) PRF:  
**SEE FIG 1 & 2.**
- f) Jitter, stability:  
**OK**
- g) Prime power:

**5 VOLTS/DIV**  
**2 NSEC/DIV**  
**400 KHZ.**

**+15 ± 1.0 V**  
**125 mA.**

*[Handwritten signature]*