


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH

S.N.: 953

Date: July 30 82



a) Output signal amplitude:

$\approx +30V$

b) Pulse width:

$\leq 1 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ VOLT}$

175 mA (max)

16 VOLTS/DIV
1.0 NSEC/DIV
400 KHZ