


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH

S.N.: 954

Date: AUG 30 82



a) Output signal amplitude:

$\geq +30V$

b) Pulse width:

$\leq 1 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ VOLTS}$

$150 \text{ mA (MAX)}$

16 VOLT/DIV

1.0 nsec/DIV

400 kHz