


PULSE GENERATOR  
PERFORMANCE CHECK

Model: **17VK-P**  
S.N.: **957**  
Date: **JULY 31 82**



- a) Output signal amplitude:  
**> +15V**
- b) Pulse width:  
**3 NSEC - 100 NSEC.**
- c) Rise time:  
**≤ 100 PSEC**
- d) Fall time:  
**≤ 2 NSEC**
- e) PRF:  
**SEE PG 1+2**
- f) Jitter, stability:  
**OK**
- g) Prime power:

**10 VOLT/DIV.**  
**1.0 NSEC/DIV**  
**200 KHZ.**

**+15 ± 1.0 V**  
**130 mA.**

**AM**