


PULSE GENERATOR

PERFORMANCE CHECK

Model: **170K-P**  
S.N.: **958**  
Date: **JULY 31 82**



- a) Output signal amplitude:  
 **$> 15$  VOLTS**
- b) Pulse width:  
**3 NSEC TO 100 NSEC**
- c) Rise time:  
 **$\leq 100$  PSEC**
- d) Fall time:  
 **$\leq 2$  NSEC**
- e) PRF:  
**SEE FILE 1+2**
- f) Jitter, stability:  
**OK.**
- g) Prime power:

**10 VOLTS/DIV.**  
**1.0 NSEC/DIV.**  
**200 KHZ.**

**$+15 \pm 1.0$  V**  
**130 mA.**